


Determination of Maltose Content


Application #AN6050

Conditions


Column: ACE Excel 5 NH₂
Dimensions: 250 x 4.6 mm
Part Number: EXL-1214-2546U
Mobile Phase: MeCN/H₂O (70:30 v/v)
Flow Rate: 1.0 mL/min
Injection: 20 µL
Temperature: 35 °C
Detection: Refractive Index
Method: ChP.2015, Vol. 4



1. Glucose



2. Maltose



3. Maltotriose

